

Product Description

th Minuteman is an intermediate level hard surface disinfectant as well as a tuberculoride virucide and fungicide, tb Minuteman is designed for use in dental and medical clinics, hospitals and nursing homes where disinfection and prevention of cross-contamination is of prime importance. the Minuteman is safe for use on non-porous environmental hard surfaces, such as counters, travs, tables, etc. tb Minuteman will not stain or leave residual smears after drying, tb Minuteman is not for use in hospital nurseries.

Product Information

tb Minuteman provides complete tuberculocidal activity against: • Mycobacterium boyis (BCG)

th Minuteman provides complete bactericidal activity against:

Pseudomonas aeruginosa

- Salmonella choleraesius
- Staphylococcus aureus
- . Methicillin Resistant
- Staphylococcus aureus (MRSA)
- Vancomycin Resistant Enterococcus faecalis (VRE)

th Minuteman provides complete fungicidal activity against: • Trichophyton mentagrophytes

- tb Minuteman provides complete virucidal activity against: • Influenza A virus
 - · Poliovirus type 1. Sabin

 - Hernes simplex virus type 1
 - Herpes simplex virus type 2
 - Human Immunodeficiency virus type 1 (HIV-1)*

Directions for Use

th Minuteman is a Ready to Use product (RTU) and should be used at full strength. Preclean the surface by liberally applying the Minuteman then wiping with a clean cloth to remove organic soil.

To disinfect the cleaned surface, apply tb Minuteman liberally, until the surface is thoroughly wet. Ensure the surface remains wet for at least 1 minute at Room Temperature. Then wipe the surface dry.

Disinfection of Non-Critical Instruments

This product is not to be used as a sterilant/high-level disinfectant on any surface or instrument that: (1) is introduced directly into the human body, either into or in contact with the bloodstream or normally sterile areas of the body, or (2) contacts intact mucous membranes but which does not ordinarily penetrate the blood barrier or otherwise enter normally sterile areas of the body.

This product may be used to pre-clean or decontaminate critical or semi-critical medical devices prior to sterilization or high-level disinfection. Instruments must first be precleaned to remove excess organic debris. then rinsed with potable water and air-dried. Immerse instruments or devices in the Minuteman for a minimum of 1 minute at Room Temperature. Remove from the solution, drain excess liquid and allow instruments to air dry. Use a fresh solution daily.

* Specific Directions for Cleaning and Decontamination Against HIV-1 on Surfaces/Objects Previously Soiled with Blood or Body Fluids

This product is intended for use against HIV-1 in settings where the virus would be expected to be encountered, such as settings where contamination by blood or body fluid is likely.

Personal Protection: Personnel that clean items soiled with blood or body fluids must wear. appropriate personal protective equipment (i.e., barrier protection), such as disposable gloves, gowns, masks or eye coverings.

Cleaning: Target surfaces must be cleaned prior to disinfection, because the presence of heavy soil on a surface may reduce the intended disinfectant efficacy of the product. Infectious Materials Disposal: Any materials used in the cleaning process that may contain blood or body fluids are to be disposed of immediately in accordance with local regulations for infectious materials disposal.

Precautions:

Refer to Directions for Use.

May cause eye irritation. Avoid contact with eyes and skin. Use in a well ventilated area. Avoid excessive inhalation of vapours. Keep out of reach of children.

First Aid:

If swallowed, give a glass of water. If contact with eyes or skin occurs, flush thoroughly with water. For excessive inhalation, remove patient to fresh air. Contact a physician if ingested, if contact with eyes occurs or if an irritation persists.

Storage:

Flammable, Keep away from heat, sparks and open flames. Store in a well ventilated environment at Room Temperature.

Disposal:

Rinse the empty container. Do not reuse the empty container. Dispose of the container in accordance with municipal and provincial requirements. For information on spills and for clean up of spills or the disposal of unused, unwanted product, consult the SDS or contact maxill or the Provincial Regulatory Agency.



Product of Canada REV 15/04