

nitriCLEAR^{AF}

POWDER FREE NITRILE MEDICAL EXAMINATION GLOVES

Standards Met:

- ASTM D6978-05
- ASTM F1670-08(2014)
- ASTM F1671:2007
- ASTM D7558-09(2014)
- ASTM D6319

Attributes:

- Not made with accelerators
- Not made with natural rubber latex
- Ambidextrous
- Short beaded cuff
- Textured fingertip
- Non-sterile
- Single use only


Dimensions:

Box dimensions: 23.5 cm x 11.8 cm x 8 cm
Box count: 200 gloves per box
Case pack: 10 boxes per case



200 GLOVES

SPECIFICATIONS	X-SMALL	SMALL	MEDIUM	LARGE	X-LARGE
Item Number	#51490	#51491	#51492	#51493	#51494
Width of Palm (mm)	75 ± 5	85 ± 5	95 ± 5	105 ± 5	115 ± 5
Total Length (mm)	230 min.	230 min.	230 min.	230 min.	230 min.
Thickness: Palm (mm)	0.05 min.	0.05 min.	0.05 min.	0.05 min.	0.05 min.
Thickness: Finger (mm)	0.05 min.	0.05 min.	0.05 min.	0.05 min.	0.05 min.
Thickness: Cuff (mm)	0.05 ± 0.01	0.05 ± 0.01	0.05 ± 0.01	0.05 ± 0.01	0.05 ± 0.01
Tensile Strength (MPa)	14 min.	14 min.	14 min.	14 min.	14 min.
Elongation (%)	Before Aging	500 min.	500 min.	500 min.	500 min.
Elongation (%)	After Aging	400 min.	400 min.	400 min.	400 min.



Made in China

The information provided is given in good faith and believed to be accurate, but no warranty or guarantee is given or implied. Conditions of use under which the product is being applied will vary from the conditions of normal laboratory efficacy testing. Therefore, each health care facility should conduct its own efficacy testing in order to verify the efficacy of the product under its own normal operating conditions to meet the specific requirements of the workplace.

MP-REF-51490B
Revision 1: 31 January 2018