Testing of maxill Silken and maxill Plus masks

	Maxill Unique Device Identifier	6070553011 6070553012 6070553013 6070553015 6070553021[1] 6070553022 6070553023 6070553024 6070553025 6070553031 6070553032 6070553033 6070553034 6070553035	6070553001 6070553002 6070553004 6070553041 6070553042 6070553043 6070553045 6070553051 6070553052 6070553053 6070553054 6070553055
	ASTM Type	Level 2	Level 3
Delta P MIL-M-36954C	Average Result	2.5 mmH2O/cm2	3.4 mmH2O/cm2
	Testing Report Reference	ITR-2015-10-15, 2015-10-22, 2015-11-02	ITR-2015-10-15, 2015-10-22, 2015-11-02
Bacterial Filtration Efficiency ASTM F2101-07	Average Result	99.70%	99.70%
	Testing Report Reference	828147	828147
Particle Filtration Efficiency ASTM F2299	Average Result	99.26%	99.26%
	Testing Report Reference	828148	828148
Synthetic Blood Penetration Resistance (SBPR) ASTM F1862	Average Result	Pass 120 mm Hg	Pass 160 mm Hg
	Testing Report Reference	00532-15	00808-15

[1.] This document corresponds only to the UDI 6070553021 and doesn't apply to UDI 6070553021F or UDI 6070553021J