maxill

35% Phosphoric Acid Dental Etch Gel



FOR PROFESSIONAL DENTAL USE

Contents

- 4 x 5 mL Gel Filled Syringes
- Dispensing Tips
- Syringe Sleeves
- Instructions for Use

Reorder No. REF 60060

maxill 35% Phosphoric Acid Dental Etch Gel

Composition: Water, phosphoric acid, glycerin, xanthan gum, sodium benzoate, colour (D&C Green #5)

Purpose/Indication: 35% Phosphoric Acid Dental Etch Gel is a surface-conditioning agent for treatment of enamel and/or dentin prior to adhesive restoration placement.

Contraindication: 35% Phosphoric Acid Dental Etch Gel is contraindicated for use on dental pulp tissue.

⚠ PRECAUTIONS RONLY

- This product is intended to be used only as specifically outlined in the **DIRECTIONS FOR USE**.
 Any use of this product inconsistent with the Directions for Use is at the discretion and responsibility of the practitioner.
- 2. Wear suitable protective eyewear, clothing and gloves. Ensure patient's eyes are protected in case of inadvertent squirting of the gel.
- 3. Replace original cap of 35% Phosphoric Acid Dental Etch Gel tightly after each use to avoid drying and clogging at the tip.
- 4. Avoid injury or cross contamination; discard dispensing tip after each use.

The tips are single use only.

- 5. Store syringe with original cap. Do not store syringe with a dispensing tip in place.
- 6. Store between 8° and 25°C away from sunlight. Prolonged excessive heat will reduce viscosity.
- 7. This gel should extrude easily. **Do not use excessive force which may lead to unanticipated extrusion of the material.**
- 8. This product should not be used on individuals who may have an allergic reaction to any of the components.
- 9. Do not use after the expiration date.
- 10. Refer to the Safety Data Sheet.

PROCEDURES FOR INFECTION CONTROL

The syringes are intended to be used multiple times and are prone to exposure to microorganisms from the oral cavity. Before each procedure encase the syringe in a new, disposable barrier sleeve. Discard sleeve after each use (single use only). Optional additional step: Disinfect the cap and/or syringe barrel using a compatible intermediate level disinfectant wipe.

DIRECTIONS FOR USE

- 1. Prepare tooth and cleanse the enamel surface to be etched. Rinse and dry thoroughly
- 2. Attach a Luer Lock dispensing tip to the syringe so that it is fully seated. **Start the flow on a pad, not in the patient's mouth. The gel should flow freely.**
- 3. Apply 35% Phosphoric Acid Dental Etch Gel to enamel and dentin surfaces and allow to contact for 15 to 20 seconds.

Never use excessive force on syringe plunger to dispense gel which may lead to over dispensing of material.

Never draw back over dispensed material into the syringe.

4. Evacuate the gel while rinsing thoroughly with copious amounts of water. It is important not to splash the gel into eyes, onto soft tissues or skin. ▶

DIRECTIONS FOR USE (CONTINUED)

- 5. Leave the surface wet or dry with clean, uncontaminated air without desiccating the dentin, depending on the instructions for use for the bonding system being used.
- 6. Do not allow etch surface to become contaminated with saliva or blood during the bonding procedure. If contamination occurs, repeat this procedure from Step 1.
- 7. Remove the dispensing tip and discard. Re-cap the syringe immediately.
- 8. Proceed with placement of the bonding agent.

GHS LABEL

Dang

May be corrosive to metals Harmful if swallowed May be harmful in contact with skin Causes severe skin burns and eye damage





Keep only in original container.

Do not breathe dusts or mists.

Wash hands thoroughly after handling.

Do not eat, drink or smoke when using this product.

Wear protective gloves/eye protection/face protection.

IF SWALLOWED: Rinse mouth. Do NOT induce vomiting. Call a POISON CENTER/doctor if you feel unwell.

IF ON SKIN: Take off immediately all contaminated clothing.

Rinse skin with water/shower. Wash contaminated dothing before reuse. Call a POISON CENTER/doctor if you feel unwell.

IF INHALED: Remove person to fresh air and keep comfortable for breathing. Immediately call a POISON CENTER/doctor.

IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER/doctor.

IF ON ORAL MUCOSA: Avoid contact with oral soft tissues. If accidental contact occurs, flush mucosa with plenty of water and expectorate water. If sensitization, inflammation or sloughing of mucosa persists, seek medical attention immediately. Immediately call a POISON CENTER/doctor.

Absorb spillage to prevent material damage.

Store locked up. Store in original container.

Dispose of contents/container in accordance with local/regional/national/international regulations.

Ouestions? Call 1-800-268-8633 or visit www.maxill.com

maxill inc.
St. Thomas, ON
N5R 6C8 Canada
www.maxill.com

maxill inc.
Cortland, OH
44410 USA

Reorder No.

Ref 60060 Revision 1: 2016-11-04