

Cleaning and Sterilization Instructions for
Reusable Cheek Retractors

(Item 41500/41501/41502/41503/41504/41505/41507/41508/41509/41510)




Extraoral Cheek Retractors

Small . . . . . #41500
Medium. . . . . .#41501
Large . . . . . #41502




Intraoral Cheek Retractors

Small . . . . . #41503
Medium. . . . . #41504
Large . . . . . #41505



Hand Held Cheek Retractors

Child. . . . . #41507
Adult . . . . . #41508



Double Hand Held Cheek Retractors

Child. . . . . #41509
Adult . . . . . .#41510

Before and after use the cheek retractors must follow a cleaning procedure and proper sterilization.

Cleaning

Cheek retractors should be cleaned with an enzymatic cleaner in an ultrasonic bath. An instrument washer can be used for the cleaning process provided the temperature does not exceed 135°C (275°F). Ensure retractors are thoroughly rinsed and fully dried prior to packaging for sterilization. Improper rinsing can leave a chemical residue from the cleaning agent and cause a reaction with the high heat from sterilization.

Sterilization

Cheek retractors can be heat sterilized in an autoclave sterilizer up to 135°C (275°F). After sterilization, cheek retractors are to remain in their sealed package and opened only at point of use.

maxill inc.®
www.maxill.com 1-800-268-8633

MP-REF-41500D
Rev.0: 2018-06-11