

nes PLUTO™

POWDER FREE NITRILE MEDICAL EXAMINATION GLOVES

Standards Met:

ASTM D6319

Attributes:

- Not made with natural rubber latex
- Powder Free
- Ambidextrous
- Short beaded cuff
- Textured fingertip
- Non-sterile
- Single use only




Dimensions:

Box dimensions: 24 cm x 12.8 cm x 8.3 cm

Box count: 300 gloves per box (XS-L), 250 gloves per box (XL)

Case pack: 10 boxes per case

SPECIFICATIONS	X-SMALL	SMALL	MEDIUM	LARGE	X-LARGE
Item Number	#51420	#51421	#51422	#51423	#51424
Width of Palm (mm)	70 - 80	80 - 90	90 - 100	100 - 110	110 - 120
Total Length (mm)	230 min.	230 min.	230 min.	230 min.	230 min.
Thickness: Cuff (mm)	0.04 min.	0.04 min.	0.04 min.	0.04 min.	0.04 min.
Thickness: Palm (mm)	0.05 min.	0.05 min.	0.05 min.	0.05 min.	0.05 min.
Thickness: Finger (mm)	0.06 min.	0.06 min.	0.06 min.	0.06 min.	0.06 min.
Tensile Strength (MPa)	14 min.	14 min.	14 min.	14 min.	14 min.
Elongation (%)	Before Aging	500 min.	500 min.	500 min.	500 min.
Elongation (%)	After Aging	400 min.	400 min.	400 min.	400 min.



Made in Malaysia

The information provided is given in good faith and believed to be accurate, but no warranty or guarantee is given or implied.

MP-REF-51420BU
 Rev.1: 2018-11-05